

Opelstr. 14
D-64546 Mörfelden-Walldorf
Telefon: +49-(0)6105-9301-0
Telefax: +49-(0)6105-9301-15
E-Mail: info@acila.com
Internet: www.acila.com

KONFORMITÄTSERKLÄRUNG CERTIFICATE of CONFORMITY

Datum der Konformität / date of conformity:	2025-12-10
Produkt / product:	PYROKONTROL® 0,1 N LAL-Natronlauge / 0,1 N LAL-Sodium hydroxide solution
Volumen / volume:	100 ml
Artikelnr. / product no.:	1712200
Ch.-B. / lot no.:	29971
Verfallsdatum / expiration date:	2029-11-30
Lagerung bei / storage at:	+2 °C bis +30 °C / +2 °C to +30 °C

D: Hiermit erklären wir, dass das o.g. Produkt den Anforderungen der EP 2.6.14. „Bacterial Endotoxins“ und der USP <85> “Bacterial Endotoxin Test” und unseren Spezifikationen entspricht.

E: We hereby certify, that the above mentioned product is in accordance with EP 2.6.14. “Bacterial Endotoxins” and USP <85> “Bacterial Endotoxin Test” and our specifications.

Endotoxin / endotoxin		Ergebnis/results
Prüfverfahren test method	LAL-Test - Chromogenic kinetic method - Ph. Eur. 2.6.14, method C - USP <85> - ANSI/AAMI ST72:2019 (Bacterial endotoxins- test methods, routine monitoring and alternatives to batch testing) - USP <1225> (Validation of Compendial Procedures) - USP <1226> (Verification of Compendial Procedures)	
Spezifikation /specification	<0,01 I.E./ml / 0.01 IU/mL	entspricht / meets

Dieses Dokument wurde elektronisch erstellt und gilt ohne Unterschrift.
The document has been generated electronically and is valid without a signature

Die Prüfergebnisse beziehen sich ausschließlich auf die Prüfgegenstände. Das Zertifikat darf auszugsweise nicht ohne schriftliche Genehmigung der ACILA AG vervielfältigt werden.

The test results refer exclusively to the items tested. The certificate must not be copied partially without the approval of ACILA AG.

13. Konformitätserklärung || Revisionsdatum: 08.05.2025

\\10.43.55.5\firmadata\ZERT-ZENTRUM\Konformitätserklärungen\2 Wasser und Puffer\NaOH\29971 NaOH, 0,1 N, HD-PE Flaschen, M 48429 Tu.docx